

PerfeCta™ SYBR® Green SuperMix, ROX™

Cat. No 95055-500 Size: 500 x 50- μ L reactions (10 x 1.25 mL)
 95055-02K 2000 x 50- μ L reactions (1 x 50 mL)

**Store at -25°C to -15°C
 protected from light**

Description

PerfeCta SYBR Green SuperMix, ROX is a 2X concentrated, ready-to-use reaction cocktail that contains all components, except primers and template for real-time quantitative PCR on Applied Biosystems 7000, 7300, 7700, 7900HT, StepOne™, or StepOnePlus™ instruments. The proprietary buffer and stabilizers have been optimized exclusively for SYBR Green I qPCR to deliver maximum PCR efficiency, sensitivity, and robust fluorescent signal. This supermix provides the highest level of specificity to reduce the occurrence or delay the detection of primer-dimer and other non-specific artifacts. Highly specific amplification is crucial to successful qPCR with SYBR Green I technology because this dye binds to and detects any dsDNA generated during amplification. A key component of this supermix is AccuStart Taq DNA polymerase, which contains monoclonal antibodies that bind to the polymerase and keep it inactive prior to the initial PCR denaturation step. Upon heat activation (2 minutes at 95°C), the antibodies denature irreversibly, releasing fully active, unmodified Taq DNA polymerase. This enables specific and efficient primer extension with the convenience of room temperature reaction assembly.

Instrument Compatibility

Different real-time PCR systems employ different strategies for the normalization of fluorescent signals and correction of well-to-well optical variations. It is critical to match the appropriate qPCR reagent to your specific instrument. PerfeCta SYBR Green SuperMix, ROX provides seamless integration on the Applied Biosystems 7000, 7300, 7700, 7900HT, StepOne, or StepOnePlus. Please consult the following table, or visit our web site at www.quantabio.com to find the optimal kit for your instrument platform.

Reagent	Cat Nos	Compatible Real-Time PCR Systems
PerfeCta SYBR Green SuperMix, ROX	95055-500, 95055-02K	Applied Biosystems 7000, 7300, 7700, 7900, 7900HT, 7900HT Fast, StepOne™, StepOnePlus™
PerfeCta SYBR Green SuperMix, Low ROX	95056-500, 95056-02K	Applied Biosystems 7500, 7500 Fast, ViiA™ 7 Stratagene MX4000™, MX3005P™, MX3000P™
PerfeCta SYBR Green SuperMix for iQ	95053-500, 95053-02K	Bio-Rad iCycler iQ®, iQ™5, MyiQ™
PerfeCta SYBR Green SuperMix	95054-500, 95054-02K	Bio-Rad CFX96™, CFX384™, Opticon™, MiniOpticon™, Chromo4™ Cepheid Smart Cycler®; Qiagen/Corbett Rotor-Gene® Eppendorf Mastercycler® ep realplex Roche Applied Science LightCycler® 480

Components

PerfeCta SYBR Green SuperMix, ROX (2X): 2X reaction buffer containing optimized concentrations of MgCl₂, dNTPs (dATP, dCTP, dGTP, dTTP), AccuStart Taq DNA Polymerase, SYBR Green I dye, ROX Reference Dye, and stabilizers.

Storage and Stability

Store components in a constant temperature freezer at -25°C to -15°C protected from light upon receipt.
 For lot specific expiry date, refer to package label, Certificate of Analysis or Product Specification Form.

Guidelines for SYBR Green qPCR:

- The design of highly specific primers is the single most important parameter for successful real-time PCR with SYBR Green I dye. The use of computer aided primer design programs is encouraged in order to minimize the potential for internal secondary structure and complementation at 3'-ends within each primer and the primer pair. PerfeCta SYBR Green SuperMix, ROX can readily amplify fragments between 400 and 500 bp; however, for best results, amplicon size should be limited to 80 - 200 bp. Optimal results may require titration of primer concentration between 100 and 500 nM. A final concentration of 300 nM for each primer is effective for most reactions.
- Preparation of a reaction cocktail is recommended to reduce pipetting errors and maximize assay precision. Assemble the reaction cocktail with all required components except sample template (genomic DNA or cDNA) and dispense equal aliquots into each reaction tube. Add the DNA template to each reaction as the final step. Addition of samples as 5 to 10- μ L volumes will improve assay precision.
- Suggested input quantities of template are: cDNA corresponding to 1 pg to 100 ng of total RNA; 100 pg to 100 ng genomic DNA
- After sealing each reaction, vortex gently to mix contents. Centrifuge briefly to collect components at the bottom of the reaction tube.



Reaction Assembly

Component	Volume for 50- μ L rxn.	Final Concentration
PerfeCta SYBR Green SuperMix, ROX (2X)	25 μ L	1x
Forward primer	variable	100 – 500 nM
Reverse primer	variable	100 – 500 nM
Nuclease-free water	variable	
Template	<u>5 – 10 μL</u>	variable
Final Volume (μ L)	50 μ L	

Note: For smaller reaction volumes (i.e. 25- μ L reactions), scale all components proportionally.

Reaction Protocol

Incubate complete reaction mix in a real-time thermal detection system as follows:

Initial denaturation:	95°C, 2 to 3 min
PCR cycling (30-45 cycles):	95°C, 10 to 15 s
	55 – 65°C, 30 to 45 s (collect and analyze data)
Melt Curve (dissociation stage)	Refer to instrument instructions (optional)

Full activation of AccuStart Taq DNA polymerase occurs within 30 seconds at 95°C. Initial denaturation times greater than 3 minutes are usually not required when amplifying cDNA template. However, amplification of genomic DNA or supercoiled plasmid DNA targets may benefit from a prolonged initial denaturation step (5-10 min) to fully denature and fragment the template. This minimizes the potential for renaturation of long fragments and/or repetitive sequence regions that can impair replication of the target sequence by the PCR process.

Some primer sets may require a 3-step cycling protocol for optimal performance. Optimal annealing temperature and time may need to be empirically determined for any given primer set. A 68 to 72°C extension step of 30 seconds is suitable for most applications. However, amplicons greater than 200 bp may require longer extension times. The use of an elevated temperature (80°C) for data collection is not recommended. While this technique can be used to mask the detection of primer-dimer and/or other non-specific products, it does little to improve assay specificity or sensitivity and is not a substitute for effective primer design.

Quality Control

Kit components are free of contaminating DNase and RNase. PerfeCta SYBR Green SuperMix, ROX is functionally tested in qPCR. Kinetic analysis must demonstrate linear resolution over six orders of dynamic range ($r^2 > 0.995$) and a PCR efficiency $> 90\%$.

Limited Label Licenses

Use of this product signifies the agreement of any purchaser or user of the product to the following terms:

1. The product may be used solely in accordance with the protocols provided with the product and this manual and for use with components contained in the kit only. Quantabio, LLC. grants no license under any of its intellectual property to use or incorporate the enclosed components of this kit with any components not included within this kit except as described in the protocols provided with the product, this manual, and additional protocols available at www.quantabio.com. Some of these additional protocols have been provided by Quantabio product users. These protocols have not been thoroughly tested or optimized by Quantabio, LLC. Quantabio, LLC neither guarantees them nor warrants that they do not infringe the rights of third-parties.
2. Other than expressly stated licenses, Quantabio, LLC. makes no warranty that this kit and/or its use(s) do not infringe the rights of third-parties.
3. This kit and its components are licensed for one-time use and may not be reused, refurbished, or resold.
4. Quantabio, LLC. specifically disclaims any other licenses, expressed or implied other than those expressly stated.
5. The purchaser and user of the kit agree not to take or permit anyone else to take any steps that could lead to or facilitate any acts prohibited above. Quantabio, LLC. may enforce the prohibitions of this Limited License Agreement in any Court, and shall recover all its investigative and Court costs, including attorney fees, in any action to enforce this Limited License Agreement or any of its intellectual property rights relating to the kit and/or its components.

©2021 Quantabio, LLC. 100 Cummings Center Suite 407J Beverly, MA 01915; Telephone number: 1-888-959-5165.

Quantabio products are manufactured in Beverly, Massachusetts, Frederick, Maryland and Hilden, Germany.

Intended for molecular biology applications. This product is not intended for the diagnosis, prevention or treatment of a disease.

PerfeCta is a registered trademark of Quantabio, LLC.

TaqMan is a registered trademark of Roche Molecular Systems, Inc. LightCycler is a registered Trademark of Roche. Applied Biosystems, StepOne, StepOnePlus, ViiA, and ROX are trademarks Life Technologies Corporation. Stratagene, MX3000P, MX3005P and MX4000 are trademarks of Agilent Technologies, Inc. Mastercycler is a trademark of Eppendorf. Rotor-Gene is a registered trademark of Qiagen GmbH. SmartCycler is a trademark of Cepheid. CFX96, CFX384, iCycler iQ, iQ5, MyiQ, Opticon, MiniOpticon and Chromo4 are trademarks of Bio-Rad Laboratories. SYBR is a registered Trademark of Molecular Probes, Inc.