# Quantabio

Product Information			
qScript <sup>®</sup> 1-Step qRT-PCR Kit ROX™			
Part Number	95058-200		
Number of Reactions	200 Reactions		
Reaction Size	50 μL		
Storage Temperature	-25ºC to -15ºC		
Lot Number	66154630		
Reference Number	020218, 051018		
Expiration Date	02/28/2021		

Component Part Numbers: 84022 qScript 1-Step RT 0.05 mL 84037 2X 1-Step Master Mix ROX 1.25 mL 84007 Nuclease Free Water 1.5 mL

## Product Description:

The qScript One-Step qRT-PCR ROX Kit is a convenient and highly sensitive solution for reverse transcription quantitative PCR (RT-qPCR) of RNA templates using hybridization probe detection chemistries such as TaqMan(r) 5'-hydrolysis probes. cDNA Synthesis and PCR amplification are carried out in the same tube without opening between procedures. The system has been optimized to deliver maximum RT-PCR efficiency, sensitivity, and specificity, enabling unbiased coamplification of low copy transcripts in the presence of higher copy reference genes. The proprietary reaction buffer has been specifically formulated to maximize activities of both reverse transcriptase and Taq DNA polymerase while minimizing the potential for primerdimer and other non-specific PCR artifacts.

Product Specifications			
95058			
Assay	qScript 1 Step qRT-PCR Functional Assay	DNase	RNase
Result	Pass	Pass	Pass

## Quality Control Analysis and Specifications:

## Nuclease Assay:

**DNase:** DNase activity must be below the detectable limits of 100 pg DNase I equivalent as assayed using a fluorogenic substrate following a 1 hour incubation at 37°C with each kit component at 1X concentration. **RNase:** RNase activity must be below the detectable limits of 1 pg RNase A equivalent as assayed using a fluorogenic substrate following a 1 hour incubation at 37°C with each kit component at 1X concentration.

**qScript 1-Step qRT-PCR Functional Assay:** Detection of ACTB gene from 1.0 pg to 1.0  $\mu$ g of UHR RNA. Ct standard curve analysis must have a coefficient of determination R<sup>2</sup>  $\ge$  0.990 with a slope between -3.20 and -3.65. No template controls below threshold for at least two replicates.

#### Limitations of Use

Quantabio and Ultraplex are registered trademarks of QIAGEN Beverly, Inc. Quanta Biosciences, qScript, Geltrack, ToughMix, PerfeCta, and Fastmix are registered trademarks of Quanta BioSciences Inc. Extracta, AccuStart, AccuMelt, and Accuvue are trademarks of Quanta BioSciences Inc. Applied Biosystems, StepOne, StepOnePlus and ROX are trademarks of Thermo Fisher Scientific and or its subsidiaries. Please contact QIAGEN-Beverly for more information.

This product was developed, manufactured, and sold for *in vitro* use only. The product is not suitable for administration to humans or animals. SDS sheets relevant to this product are available upon request. 100 Cummings Center, Suite 407J, Beverly, MA 01915 • Ph (888) 927-7027 • Fax (978) 867-5724 • www.QuantaBio.com FMWI016.2 Rev 01