



Product Information				
qScript® MicroRNA cDNA Synthesis				
Part Number	95107-025			
Number of Reactions	25 Reactions			
Reaction Size	20 μL			
Storage Temperature	-25°C to -15°C			
Lot Number	026340			
Reference Number	012517, 061217, 051717, 031317, 160550828, 191716390			
Expiration Date	03/31/2020			

**Product Description:** 

The qScript microRNA cDNA Synthesis Kit is an optimized reagent system that reverse transcribes small single-stranded RNA into 5'-labeled cDNA using either total RNA or miRNA enriched samples.

Single-stranded RNA is first polyadenylated by poly(A) polymerase before reverse transcription into universal cDNA using high performance *qScript* RT with a proprietary adapter oligo(dT) primer. The universal cDNA template enables simple and cost-effective microRNA profiling when used together with wet-lab validated PerfeCTa microRNA Assays, PerfeCTa Universal PCR Primer and PerfeCTa SYBR Green SuperMix.

Component Part Numbers: 84001 qScript RT 0.025 mL

84172 Poly (A) Tailing Buffer (5X)
84174 Poly (A) Polymerase
84176 MicroRNA cDNA Reaction Mix
84178 PerfeCTa Universal PCR Primer
84180 PerfeCTa Human Positive Control Primer
84007 Nuclease Filled Water 1.50 mL

Product Specifications				
95107				
Assay	qScript microRNA cDNA Synthesis Kit	β-actin SYBR Green qRT-PCR Assay for qScript Reverse Transcriptase	DNase	RNase
Result	Pass	Pass	Pass	Pass

## **Quality Control Analysis and Specifications:**

## **Nuclease Assay**

**DNase:** DNase activity must be below the detectable limits of 100 pg DNase I equivalent as assayed using a fluorogenic substrate following a 1 hour incubation at 37°C with each kit component at 1X concentration.

**RNase:** RNase activity must be below the detectable limits of 1 pg RNase A equivalent as assayed using a fluorogenic substrate following a 1 hour incubation at 37°C with each kit component at 1X concentration.

**qScript microRNA cDNA Synthesis Kit** Must have detection of microRNA from 100 ng to 10 pg. Correlation of determination  $(R^2) \ge 0.990$  from Ct standard curve analysis. Slope from Ct standard curve analysis between -3.20 and -3.75. The +PAP reactions must show a  $\ge 11$  Ct difference from the "no-PAP" reactions. There must be at least three negative template controls with "undetermined" values.

**β-actin SYBR Green qRT-PCR Assay for qScript Reverse Transcriptase:** Detection of β-actin mRNA from 100ng to 100fg of total RNA. Coefficient of determination ( $R^2$ )  $\geq 0.990$  with a slope analysis between -3.20 and -3.70

## Limitations of Use

Quantabio and Ultraplex are registered trademarks of QIAGEN Beverly, Inc. Quanta Biosciences, qScript, Geltrack, ToughMix, PerfeCta, and Fastmix are registered trademarks of Quanta BioSciences Inc. Extracta, AccuStart, AccuMelt, and Accuvue are trademarks of Quanta BioSciences Inc. Applied Biosystems, StepOne, StepOnePlus and ROX are trademarks of Thermo Fisher Scientific and or its subsidiaries. Please contact QIAGEN-Beverly for more information.

This product was developed, manufactured, and sold for *in vitro* use only. The product is not suitable for administration to humans or animals. SDS sheets relevant to this product are available upon request.

100 Cummings Center, Suite 407J, Beverly, MA 01915 • Ph (888) 927-7027 • Fax (978) 867-5724 • www.QuantaBio.com FMWI016.2 Rev 01