

Product Information				
AccuStart™ Genotyping ToughMix® Low ROX®				
Part Number	95117-012			
Number of Reactions	1250 reactions			
Reaction Size	20μL			
Storage Temperature	-25°C to -15°C			
Lot Number	027883			
Reference Number	010417			
Expiration Date	09/30/2019			

Product Description:

AccuStart Genotyping ToughMix Low ROX is a 2X concentrated ready-to-use reaction cocktail for PCR amplification of DNA templates that overcomes many known inhibitors of PCR often present in crude samples extracted from environmental specimens, plant tissues, or animal tissues. It is a versatile and robust real-time qPCR reagent that provides maximum sensitivity and PCR efficiency with a variety of fluorogenic probe chemistries, including TaqMan hydrolysis probes. The only user-supplied components are primers, probe(s), and DNA template. Inert AccuVue plate loading dye is compatible with either white or clear PCR plates and helps to minimize pipette error and provides visual confirmation of thorough mixing.

This proprietary formulation has been rigorously optimized to destabilize probes that differ by only a single base-pair substitution to provide precise allelic discrimination for genotyping single nucleotide polymorphisms (SNPs) with either hydrolysis (TaqMan) or FRET-based fluorescent probes.

Component Part Numbers:

84206 AccuStart Genotyping ToughMix Low ROX, 1.25mL

Product Specifications					
	95117				
Assay	RT-qPCR Plasmid DNA Functional Assay	RT-qPCR genomic DNA Functional Assay	DNase	RNase	
Result	Pass	Pass	Pass	Pass	

Quality Control Analysis and Specifications:

Nuclease Assay:

DNase: DNase activity must be below the detectable limits of 100 pg DNase I equivalent as assayed using a fluorogenic substrate following a 1 hour incubation at 37°C with each kit component at 1X concentration.

RNase: RNase activity must be below the detectable limits of 1 pg RNase A equivalent as assayed using a fluorogenic substrate following a 1 hour incubation at 37°C with each kit component at 1X concentration.

RT-qPCR Plasmid DNA Functional Assay: Fast-cycling Real-time PCR detection of log-fold serial dilutions of a control DNA from 10 to 1×10^7 copies. Linear regression analysis of cycle threshold versus log input quantity must give a slope of between -3.20 and -3.65 and coefficient of determination (R^2) ≥ 0.990 .

RT-qPCR genomic DNA Functional Assay: Real-time PCR detection of single-copy gene in human genomic DNA using activation step of 10 minutes at 95°C. Linear regression analysis of cycle threshold versus log input quantity for a log-fold serial dilutions of human genomic DNA from 10 to 1 x 10^5 copies must give a slope of between -3.20 and -3.65 and coefficient of determination (R^2) ≥ 0.990 with accurate two-fold discrimination of 500, 1000, and 2000 copies.

Limitations of Use

Quantabio and Ultraplex are registered trademarks of QIAGEN Beverly, Inc. Quanta Biosciences, qScript, Geltrack, ToughMix, PerfeCta, and Fastmix are registered trademarks of Quanta BioSciences Inc. Extracta, AccuStart, AccuMelt, and Accuvue are trademarks of Quanta BioSciences Inc. Applied Biosystems, StepOne, StepOnePlus and ROX are trademarks of Thermo Fisher Scientific and or its subsidiaries. Please contact QIAGEN-Beverly for more information.

This product was developed, manufactured, and sold for *in vitro* use only. The product is not suitable for administration to humans or animals. SDS sheets relevant to this product are available upon request.

100 Cummings Center, Suite 407J, Beverly, MA 01915 • Ph (888) 927-7027 • Fax (978) 867-5724 • www.QuantaBio.com FMWI016.2 Rev 01