# Quantabio

Product Information				
PerfeCTa <sup>®</sup> FastMix II, Low ROX™				
95120-012				
1250 reactions				
20µL				
-25ºC to -15ºC				
023142				
092217				
08/31/2020				

## Product Description:

PerfeCTa FastMix II, Low ROX is an advanced 2X concentrated qPCR SuperMix reagent for both fast and conventional PCR cycling protocols or instruments. It is a versatile and robust solution that provides the ultimate sensitivity and high PCR efficiency using a variety of fluorogenic probe chemistries, including TaqMan hydrolysis probes. The kit is provided as a 2X concentrated ready-to-use reaction cocktail that contains all required reaction components, except primers, probe(s), and DNA template. Inert AccuVue plate loading dye helps to minimize pipette error and provides visual confirmation of thorough mixing.

A key component of the kit is an ultra pure, processive thermostable DNA polymerase that is free of detectable E. coli DNA. PerfeCTa FastMix II is ideal for demanding qPCR applications such as bacterial pathogen detection

where residual host DNA in typical recombinant enzyme preparations can limit assay sensitivity and obscure detection of low copy samples. Formulated for maximum assay precision with an ultra pure, high avidity monoclonal antibody for stringent suppression for reaction assembly at ambient room temperature and complete activity release and minimal activation time (1-3 minutes).

## Component Part Numbers: 84212 Perfecta FastMix II, Low ROX 1.25mL

Product Specifications					
95120					
Assay	RT-qPCR Plasmid DNA Functional Assay	RT-qPCR genomic DNA Functional Assay	DNase	RNase	
Result	Pass	Pass	Pass	Pass	

## **Quality Control Analysis and Specifications:**

## Nuclease Assay:

**DNase:** DNase activity must be below the detectable limits of 100 pg DNase I equivalent as assayed using a fluorogenic substrate following a 1 hour incubation at 37°C with each kit component at 1X concentration.

**RNase:** RNase activity must be below the detectable limits of 1 pg RNase A equivalent as assayed using a fluorogenic substrate following a 1 hour incubation at 37°C with each kit component at 1X concentration.

**RT-qPCR Plasmid DNA Functional Assay:** Fast-cycling Real-time PCR detection of log-fold serial dilutions of a control DNA from  $10^7$  to 10 copies. Linear regression analysis of cycle threshold versus log input quantity must give a slope of between -3.2 and -3.6 and coefficient of determination ( $r^2$ )  $\geq 0.990$ .

**RT-qPCR genomic DNA Functional Assay:** Real-time PCR detection of single-copy gene in human genomic DNA using activation step of 10 minutes at 95°C. Linear regression analysis of cycle threshold versus log input quantity for a log-fold serial dilutions of human genomic DNA from  $10^5$  to 10 copies must give a slope of between -3.2 and -3.6 and coefficient of determination ( $r^2$ )  $\geq 0.990$  with accurate two-fold discrimination of 500, 1000, and 2000 copies.

#### Limitations of Use

Quantabio and Ultraplex are registered trademarks of QIAGEN Beverly, Inc. Quanta Biosciences, qScript, Geltrack, ToughMix, PerfeCta, and Fastmix are registered trademarks of Quanta BioSciences Inc. Extracta, AccuStart, AccuMelt, and Accuvue are trademarks of Quanta BioSciences Inc. Applied Biosystems, StepOne, StepOnePlus and ROX are trademarks of Thermo Fisher Scientific and or its subsidiaries. Please contact QIAGEN-Beverly for more information.

This product was developed, manufactured, and sold for *in vitro* use only. The product is not suitable for administration to humans or animals. SDS sheets relevant to this product are available upon request. 100 Cummings Center, Suite 407J, Beverly, MA 01915 • Ph (888) 927-7027 • Fax (978) 867-5724 • <u>www.QuantaBio.com</u> FMWI016.2 Rev 01