# Quantabio

# Product Specifications 95131-500 Rev 01

| Product Information                |                |  |  |
|------------------------------------|----------------|--|--|
| qScript 1-Step Virus ToughMix (2X) |                |  |  |
| Part Number                        | 95131-500      |  |  |
| Number of Reactions                | 500            |  |  |
| Reaction Size                      | 20µL           |  |  |
| Storage Temperature                | -25°C to -15°C |  |  |
| Lot Number                         | 66171120       |  |  |
| Reference Number                   | 102820         |  |  |
| Expiration Date                    | 11/30/2021     |  |  |

#### Product Description:

qScript 1-Step Virus ToughMix is a 2x, ready-to-use, master mix for rapid detection of RNA viruses such as Flu-A, Flu-B, and SARS-CoV-2, using one-step, or single-tube reverse transcription quantitative PCR (RT-qPCR). It has been optimized for maximum sensitivity to enable reliable quantification of very low input quantities of RNA using dual-labeled hydrolysis probe detection chemistries such as TaqMan<sup>®</sup> probes in single or multiplexed assay formats. qScript 1-Step Virus ToughMix contains all required components for RT-qPCR except RNA template and primer/probe assays.

qScript 1-Step Virus ToughMix does not contain an internal reference dye.

#### Kit Components:

#### 84649 qScript Virus 1-Step Virus ToughMix, 1mL

| Product Specifications |                                 |       |       |
|------------------------|---------------------------------|-------|-------|
| 95131                  |                                 |       |       |
| Assay                  | 1-Step RT-qPCR Functional Assay | DNase | RNase |
| Result                 | Pass                            | Pass  | Pass  |

## **Quality Control Analysis and Specifications:**

**1-Step RT-qPCR Functional Assay:** Cq standard curve analysis must have coefficient of determination ( $R^2$ )  $\ge$ 0.990 with a slope between -3.20 and -3.65. Control reactions lacking template RNA (NTC) must remain below fluorescence threshold through 45 PCR cycles.

## Nuclease Assay:

**RNase:** RNase activity must be below the detectable limits of 1 pg RNase A equivalent as assayed using a fluorogenic substrate following a 1 hour incubation at 37°C with each kit component at 1X concentration. **DNase:** DNase activity must be below the detectable limits of 100 pg DNase I equivalent as assayed using a fluorogenic substrate following a 1 hour incubation at 37°C with each kit component at 1X concentration.

#### Limitations of Use

Quantabio and Ultraplex are registered trademarks of QIAGEN Beverly, Inc. Quanta Biosciences, qScript, Geltrack, ToughMix, PerfeCta, and Fastmix are registered trademarks of Quanta BioSciences Inc. Extracta, AccuStart, AccuMelt, and Accuvue are trademarks of Quanta BioSciences Inc. Applied Biosystems, StepOne, StepOnePlus and ROX are trademarks of Thermo Fisher Scientific and or its subsidiaries. Please contact QIAGEN-Beverly for more information.

This product was developed, manufactured, and sold for *in vitro* use only. The product is not suitable for administration to humans or animals. SDS sheets relevant to this product are available upon request. 100 Cummings Center, Suite 407J, Beverly, MA 01915 • Ph (888) 927-7027 • Fax (978) 867-5724 • <u>www.QuantaBio.com</u> FMWI016.2 Rev 01