# Quantabio

# Product Specifications 95166-01K Rev 01

Product Information		
UltraPlex 1-Step ToughMix®		
Part Number	95166-01K	
Number of Reactions	1000 Reactions	
Reaction Size	20 µL	
Storage Temperature	-25ºC to -15ºC	
Lot Number	66165319	
Reference Number	080720	
Expiration Date	09/30/2021	

#### Product Description:

UltraPlex 1-Step ToughMix is a ready-to-use, 4X concentrated master mix 1-step reverse transcription and real-time quantitative PCR (RT-qPCR) of RNA templates using probe-based detection methods. First-strand cDNA synthesis and PCR amplification are carried out in the same tube without opening between steps. Optimized to deliver highly sensitive quantification of RNA viruses or low abundance RNA targets in uni- or highly multiplexed RNA detection assays, this reagent chemistry is optimized to deliver maximum assay sensitivity, precision and reproducibility with miniaturized reaction volumes and either conventional or accelerated thermal cycling protocols. UltraPlex 1-Step ToughMix contains all required components for RT-qPCR except RNA template and probe and is compatible with all duallabeled probe chemistries.

Product Specifications				
	95166			
Assay	UltraPlex 1-Step ToughMix Functional Assay	DNase	RNase	
Result	Pass	Pass	Pass	

## **Quality Control Analysis and Specifications:**

### Nuclease Assay:

**DNase:** DNase activity must be below the detectable limits of 100 pg DNase I equivalent as assayed using a fluorogenic substrate following a 1 hour incubation at 37°C with each kit component at 1X concentration. **RNase:** RNase activity must be below the detectable limits of 1 pg RNase A equivalent as assayed using a fluorogenic substrate following a 1 hour incubation at 37°C with each kit component at 1X concentration.

**UltraPlex 1-Step ToughMix Functional Assay:** Must have detection of  $\beta$  actin, PPIA, GAPDH, and TUBA from 1 pg to 100,000 pg of UHR total RNA. Cq standard curve analysis for each targeted sequence must have a coefficient of determination (R<sup>2</sup>)  $\geq$ 0.990 with a slope between -3.20 and -3.65.

#### Limitations of Use

Quantabio and Ultraplex are registered trademarks of QIAGEN Beverly, Inc. Quanta Biosciences, qScript, Geltrack, ToughMix, PerfeCta, and Fastmix are registered trademarks of Quanta BioSciences Inc. Extracta, AccuStart, AccuMelt, and Accuvue are trademarks of Quanta BioSciences Inc. Applied Biosystems, StepOne, StepOnePlus and ROX are trademarks of Thermo Fisher Scientific and or its subsidiaries. Please contact QIAGEN-Beverly for more information.

This product was developed, manufactured, and sold for *in vitro* use only. The product is not suitable for administration to humans or animals. SDS sheets relevant to this product are available upon request. 100 Cummings Center, Suite 407J, Beverly, MA 01915 • Ph (888) 927-7027 • Fax (978) 867-5724 • <u>www.QuantaBio.com</u> FMWI016.2 Rev 01